

IoT based Smart Health Monitoring and Management with Cloud-based Processing

Mahendra Kumar¹, Surender Kumar²

^{1,2}Guru Kashi University, Talwandi Sabo

Abstract- Among the panoply of utilizations empowered by the Web of Things (IoT), savvy and associated social insurance is a especially vital one. Organized sensors, either worn on the body or installed in our living surroundings, make conceivable the social occasion of rich data demonstrative of our physical and emotional wellness. Caught consistently, totaled, and viably mined, such data can realize a positive transformative change in the medicinal services scene. Specifically, the accessibility of information at until now unheard of scales and fleeting longitudes combined with another age of astute handling calculations can: (an) encourage an advancement in the routine with regards to medication, from the present post facto analyze and treat responsive worldview, to a proactive system for visualization of ailments at an early stage, combined with aversion, fix, also, by and large administration of wellbeing rather than malady, (b) empower personalization of treatment and the board alternatives focused on especially to the explicit conditions and needs of the individual, and (c) help diminish the expense of medicinal services while at the same time enhancing results. In this paper, we feature the chances and difficulties for IoT in understanding this vision of things to come of medicinal services.

Keywords- remote health monitoring; IoT; visualization; analytics;

I. INTRODUCTION

Ongoing years have seen a rising enthusiasm for wearable sensors what's more, today a few gadgets are financially accessible [1]– [3] for individual social insurance, wellness, and action mindfulness. In expansion to the specialty recreational wellness field took into account by momentum gadgets, scientists have likewise thought about applications of such advancements in clinical applications in remote wellbeing observing frameworks for long haul recording, the executives and clinical access to patient's physiological data [4]– [8]. In view of current innovative patterns, one can promptly envision a period sooner rather than later when your routine physical examination is gone before by a two– multi day time of ceaseless physiological observing utilizing modest wearable sensors. Over this interim, the sensors would ceaselessly record signals corresponded with your key physiological parameters and hand-off the subsequent information to a database connected with your wellbeing records. When you appear for your physical examination, the doctor has available not only conventional clinic/lab-test based static measurements of your physiological and metabolic state, but also the much richer longitudinal record provided by the sensors. Using the available data, and aided by decision support systems that also have access to a large corpus of observation data for other individuals, the

doctor can make a much better prognosis for your health and recommend treatment, early intervention, and lifestyle choices that are particularly effective in improving the quality of your health. Such a disruptive technology could have a transformative impact on global healthcare systems and drastically reduce healthcare costs and improve speed and accuracy for diagnoses.

Technologically, the vision presented in the preceding paragraph has been feasible for a few years now. Yet, wearable sensors have, thus far, had little influence on the current clinical practice of medicine. In this paper, we focus particularly on the clinical arena and examine the opportunities afforded by available and upcoming technologies and the challenges that must be addressed in order to allow integration of these into the practice of medicine. The paper is organized as follows: Section II highlights some of the key related work in this area. In Section III, we outline the architecture for remote health monitoring systems based on wearable sensors, partitioning the system into four main components: acquisition, analytics, and visualization. In Sections IV– VII we highlight the opportunities and challenges related to each of these components. We conclude the paper in Section VIII with a summary and discussion.

II. BACKGROUND

Most proposed systems for remote wellbeing checking use a three level engineering: a Wireless Body Area System (WBAN) comprising of wearable sensors as the information obtaining unit, correspondence and organizing and the administration layer [4], [7]– [10]. For example [11] proposes a framework that initiates wearable sensors to quantify different physiological parameters, for example, circulatory strain and body temperature.

Sensors transmit the assembled data to a door server through a Bluetooth association. The entryway server turns the information into an Observation and Measurement record and stores it on a remote server for later recovery by clinicians through the Web. Using a comparative cloud based therapeutic information stockpiling, a wellbeing observing framework is exhibited in [12] in which medicinal staff can get to the put away information online through substance benefit application. Focusing on an explicit restorative application, WANDA [13] a conclusion to end remote wellbeing checking and examination framework is exhibited for supervision of patients with high danger of heart disappointment.

Notwithstanding the innovation for information gathering, stockpiling what's more, get to, therapeutic information investigation and representation are basic segments of remote wellbeing checking frameworks. Exact determinations and checking of patient's therapeutic condition depends on investigation of therapeutic records containing different physiological qualities over a significant lot of time. Managing with information of high dimensionality in both time and amount makes information investigation undertaking very disappointing and blunder inclined for clinicians. Despite the fact that the utilization of information mining and perception systems had recently been tended to as an answer for the previously mentioned test [14], [15], these techniques have recently picked up consideration in remote wellbeing checking frameworks [16], [17].

While the approach of electronic remote wellbeing checking frameworks has guaranteed to alter the customary wellbeing care strategies, incorporating the IoT worldview into these frameworks can additionally expand insight, adaptability and interoperability [9], [18]. A gadget using the IoT plot is remarkably tended to and

recognizable whenever and anyplace through the Internet. IoT based gadgets in remote wellbeing checking frameworks are not just equipped for the ordinary detecting errands however can likewise trade data with one another, naturally associate with and trade data with wellbeing founds through the Internet, altogether streamlining set up what's more, organization undertakings. As exemplified in [19], such frameworks can give administrations, for example, programmed alert to the closest medicinal services foundation in case of a basic mishap for a managed patient.

III.SYSTEM ARCHITECTURE

Figure 1 delineates the framework design for a remote wellbeing checking framework, whose significant parts we depict next: Information Acquisition is performed by various wearable sensors that measure physiological biomarkers, for example, ECG, skin temperature, respiratory rate, EMG muscle movement, and stride (act). The sensors interface with the system however an middle of the road information aggregator or concentrator, which is commonly an advanced mobile phone situated in the region of the patient.




Fig. 1. Components of a remote patient monitoring system that is based on an IoT-Cloud architecture.

The Data Transmission segments of the framework are in charge of passing on accounts of the patient from the patient's home (or any remote area) to the server farm of the Healthcare Organization (HCO) with guaranteed security furthermore, security, preferably in close continuous. Regularly, the tangible securing stage is outfitted with a short range radio as Zigbee or low-control Bluetooth, which it uses to exchange sensor information to the concentrator. Collected information is further transferred to a HCO for long haul stockpiling utilizing Internet network on the concentrator, commonly by means of a cell phone's WiFi or cell information association. Sensors in the information procurement part frame an Internet of Things (IoT)- based design as each singular sensor's information can be gotten to through the Internet through the concentrator [20], [21].

Frequently a capacity/preparing gadget in region of a versatile customer, in some cases alluded to as a cloudlet, is utilized to enlarge its stockpiling/handling capacity at whatever point the neighborhood portable assets don't satisfy the application's necessities [22]. The cloudlet can be a nearby handling unit, (for example, a work area PC) which is specifically open by the concentrator through WiFi organizes. Notwithstanding giving brief

capacity preceding correspondence of information to the cloud, the cloudlet can likewise be utilized for running time basic undertakings on the patient's collected information. Additionally, the cloudlet can be used to transmit the collected information to the cloud in the event of confinements on the cell phone, for example, impermanent absence of availability or vitality. Cloud Processing has three unmistakable parts: stockpiling, investigation, and perception. The framework is intended for long term stockpiling of patient's biomedical data too helping wellbeing experts with indicative data. Cloud based medicinal information stockpiling and the forthright difficulties have been widely tended to in the writing [23], [24]. Investigation that utilization the sensor information alongside e-Health records that are getting to be common can help with determinations and guesses for various wellbeing conditions and infections. Furthermore, Perception is a key necessity for any such framework in light of the fact that it is unrealistic to request that doctors pore over the voluminous information or examinations from wearable sensors. Perception strategies that make the information and examinations available to them in a promptly edible organization are fundamental if the wearable sensors are to affect clinical practice.

In the accompanying segments, we think about the key components

Bio-marker	CVD	COPD	PD/HD	Diabetes
Gait (posture)	✓✓	✓✓	✓✓	?
ECG	✓✓	✓✓	✓	✓
Respiratory rate	✓✓	✓✓	✓	?
Skin temperature	✓	✓	✓	✓
Surface EMG	✓	✓	✓	?
Sweating	?	?	✓	?
Blood pressure	✓✓	✓	✓	✓
Body movements	✓	?	✓✓	?
Blood Glucose	?	?	?	✓✓
Heart Sound	✓	✓	?	?
Oxygenation	✓✓	✓✓	?	?
Title volume	✓✓	✓✓	✓	?

TABLE I:List of available (top) and future (bottom) sensors and their applicability to detecting health conditions related to three common disease categories: cardiovascular diseases(cvd) [25], chronic obstructive pulmonary disease (copd) [26], and parkinson's/huntington's diseases (pd) [27], [28]. __indicates high applicability, _indicates some applicability, and? indicates undetermined applicability.of the overall system illustrated in Fig. 1 and highlight the opportunities and challenges for each in integrating remote health monitoring into clinical practice.

IV.DATA ACQUISITION AND SENSING

Physiological information is gained by wearable gadgets that consolidate small scale sensors fit for estimating different physiological parameters, minor preprocessing equipment and a correspondences stage for transmitting the deliberate information. Table I condenses different biomarkers that can be estimated by current or destined to-be-accessible wearable sensors. The dimension of appropriateness of these biomarkers to diagnosing four normal malady classes is likewise demonstrated in the table.

The wearability necessity, presents physical impediments on the plan of the sensors. The sensors must be light, little, and ought not block a patient's developments and portability. Moreover, since they have to work on little batteries incorporated into the wearable bundle, they should be vitality effective. Despite the fact that the battery might be battery-powered or replaceable, for comfort furthermore, to guarantee that information isn't lost amid energizing or battery substitution periods, it is exceedingly alluring that they give broadened lengths of nonstop activity without requiring charging or substitution.

The low vitality task prerequisite can likewise represent a challenge for the nature of the information caught as far as the reachable flag to clamor proportion. Ongoing plans [5], [29], [30] of adaptable sensors that can be set in contact with the skin in various body parts are especially alluring for medicinal applications on the grounds that, contrasted with choices, the close contact with the skin permits estimation of increasingly physiological parameters and with more noteworthy exactness. There have moreover been endeavors to delay the operational lifetime of wearable sensors by joining low power gadget and circuit level procedures [31], [32] and vitality reaping strategies [33].

In addition, using wise detecting strategies on framework level can additionally expand the operational life span. Vitality proficient detecting components have been examined in the related setting of remote sensor systems (WSNs) that are utilized to detect physical marvel in a circulated fashion. Although the sensor deployment in our health monitoring system is more concentrated compared to WSNs, existing methods for WSNs can be revisited to suit our needs. The proposed energy efficient sensing approaches revolve around assigning sensing tasks to the nodes based on their relative distance so as to sense the maximum amount of physical information while minimizing the energy consumption by removing possible redundant sensing tasks [34], [35] and by allocation of tasks based on the energy availability at each sensor [36]–[40]. Similar mechanisms can be applied to our system by defining and using a dynamic context that is based on energy availability and the health condition of the patient. For example, as indicated in Table I, individually sensed biomarkers have different levels of applicability for specific health conditions. When energy is severely limited and the vulnerable condition of the patient mandates focus on a specific biomarker, the other sensors be powered off in order to extend the lifetime. An IoT based sensing architecture facilitates the implementation of such schemes for improving energy efficiency adaptively by allowing dynamic utilization of sensors based on the context. In conventional data acquisition systems where sensors passively transmit the gathered information, such intelligence and flexibility may not be achievable.

Also by offloading the decision making process for sensing task assignment to the cloud, more sophisticated algorithms can be applied without requiring manual intervention by the patient to manipulate the sensors or the software on the data concentrator.

Energy limitation of these devices necessitates the use of suitable low power communication protocols, as the communication can account a significant part of the power consumption in sensing devices. ZigBee over IEEE 802.15.4 is commonly used in low rate WPANs (LR-WPANs) to support communication between low power devices that operate in personal operating space (POS) of approximately 10m [41]. ZigBee provides reliable mesh networking with extended battery life.

Bluetooth low energy (BLE) is another wireless communication protocol suitable for low power short range communications suitable for the unique requirements of applications such as health monitoring, sports, and home entertainment. The original Bluetooth protocol (IEEE 802.15.1) was designed to support relatively short range communication for applications of a streaming nature, such as audio. BLE modifies the framework by utilizing much longer sleep intervals to decrease the overall energy consumption. BLE achieves higher energy efficiency in terms of number of bytes sent per Joule of energy [42]. When using the aforementioned communication protocols, an intermediate node (data concentrator) is necessary to make sensors data and control accessible through Internet. To further realize the IoT concept, IPv6 over Low Power Wireless Personal Area Networks (6LoWPAN) has been proposed to seamlessly connect energy constrained WPAN devices to the Internet [19]. 6LoWPAN defines fragmentation techniques to fit IPv6 datagrams into IEEE 802.15.4 limited frame size to provide IP access to low power, low complexity sensing devices.

V. CLOUD DATA STORAGE AND PROCESSING

Information amassed by the concentrator should be exchanged to the cloud for long haul stockpiling. Offloading information stockpiling to the cloud offers advantages of adaptability and availability on interest, both by patients and clinical organizations. Moreover, used with examination and perception (portrayed in ensuing segments), cloud facilitating and preparing can decrease costs at HCOs and give better symptomatic data. In this area, we framework such could models and talk about issues that affect long haul restorative information stockpiling on the cloud.

a) Hybrid Cloud/Cloudlet Architecture: Cloudlet is a constrained asset processing and capacity stage that takes out the need to re-appropriate asset concentrated assignments to the undertaking cloud [43]– [45]. Cloudlet registering has been presented as a potential answer for convey low inactivity to time basic assignments for wellbeing checking applications by means of body region systems [46]. Correspondence among concentrator and cloudlet is acknowledged through WiFi interface. Coordinate association between these two elements decreases information exchange dormancy for time basic assignments on totaled information. LTE get to gave in concentrator can thus be utilized for direct information exchange from the concentrator to the cloud bypassing the cloudlet, while presenting the information to the dormancy forced by versatile system.

b) Context-Aware Concentration by means of Smart Devices: As recently showed, advanced cells can at as concentrators in IoT foundation as the present PDAs can utilize both LTE what's more, WiFi as the backhaul organize. Information collection can be continued either in cloudlet (exhaustive the WiFi association among concentrator and the cloudlet) or the cloud (LTE). In examines, the previous contrasted and the last mentioned, has been appeared give multiple times the throughput and to require just a tenth of the entrance time, and a large portion of the power [47], [48]. Totaled information, in any case, should be at long last be put away in the cloud to permit dispersed access and solid stockpiling. To adequately segment information collection undertakings among cloud and cloudlet, setting mindful fixation might be used. Setting can represent the present and anticipated status of the patient. For instance, at the point when the patient is in a basic condition requiring time basic observing of biosensor information, information focus might be the favored decision.

c) Privacy of the Data Concentrator: Although by and by recognizable data can be evacuated before transmitting detected information data, the framework is as yet inclined to total revelation assaults that can surmise data by means of example acknowledgment approaches [49]. Setting mindful information focus, while offering a few advantages, may likewise make detected data defenseless against total revelation assaults by enabling gatecrasher to surmise a patient's wellbeing data through system traffic examination from concentrator to versatile back pull. Standard encryption strategies can be utilized to guarantee security in such settings [50].

d) Secure Data Storage in the Cloud: Privacy is of huge significance while putting away person's electronic restorative records on the cloud. As indicated by the terms characterized by Health Insurance Portability and Accountability Act (HIPAA), the secret piece of therapeutic records must be shielded from divulgence. At the point when the therapeutic records are redistributed to the cloud for capacity, proper security protecting measures should be taken to keep unapproved parties from getting to the data. Secure distributed storage systems have accordingly been proposed for use with delicate restorative records [51]– [53]. Secure therapeutic information preparing on the cloud remains a test.

VI. ANALYTICS

Contrasted and the lab and office based estimations that are the workhorses of current clinical therapeutic practice, wearable sensors can promptly join different physiological estimations and empower social occasion of information with much better transient inspecting over any longer longitudinal time scales. These rich datasets speak to an enormous chance for information examination: machine learning calculations can conceivably perceive connections between's sensor perceptions and clinical analyses, and by utilizing these datasets over longer spans of time and by pooling over a substantial client base, enhance therapeutic diagnostics. As in other huge information applications, the utilization of investigation here can enhance precision, permit prior location, empower personalization, and diminish cost by decreasing costly lab techniques that are superfluous.

Examination on wearable sensor information can theoretically use a wide-scope of example acknowledgment and machine learning methods [54], that have developed fundamentally and are presently ordinarily accessible as tool compartments in a few programming bundles [55], [56]. A few difficulties must, be that as it may, be survived before investigation can be sent on any important scale. A portion of these difficulties are closely resembling excessively those in other huge information issues where as others are extraordinary to our setting. There are additionally, in any case, challenges extraordinary to investigation with wearable sensor information and to the restorative and clinical seeing that we're engaged upon. We feature a couple of these:

- Firstly, customary therapeutic instrumentation develops at a genuinely moderate pace. New gear and estimating gadgets normally require administrative endorsement and preparing of therapeutic staff, which restrains the rate at which new developments can be presented. The pace of advancement in hardware, then again, is a lot quicker and directed by financial contemplations to nearly line up with the purported Moore's law. This infers the wearable sensors speak to a substantially more powerfully developing set of estimation gadgets than ordinary restorative instrumentation and as new sensor modules are included, sensors are refreshed, or obsoleted, there is probably going to be a heterogeneous blend in arrangement anytime.

Machine learning strategies require further advancement to manage such heterogeneous and continually developing tactile information sources. Investigation on information from wearable sensors needs to adapt with spilling information, unavoidably missing information esteems, and information of changing dimensionality and semantics as sensor structures change after some time. Learning undertakings confront a huge challenges working in these condition, even in spite of the fact that a few advances have been made around therewith the ongoing development of Big Data applications with gigantic volumes of high-dimensional perceptions that are regularly accessible in a spilling mode. Consecutive calculations focusing on online help vector machines (SVMs) have been created both in base [57] and also double spaces [58], [59]. These calculations, notwithstanding, are most certainly not intended to manage time-changing component dimensionality, deficient information vectors because of missing highlights or obtaining disappointments, which if not treated legitimately, an truly hinder grouping execution. To adapt to missing info information, it is conceivable to credit the missing values utilizing direct or nonlinear elements of the accessible highlights, and after that continue with extrasensory learning plans dependent on full information. Ascription plans, for example, substitution with zero, passage astute mean, and weighted averaging of the K-closest neighbors are prominent interpolators, alongside other progressively refined yet in addition more exorbitant ones that can be found in [60]– [62].

- Secondly, while sensor information is ample, it is totally untagged and should be related with comparing "ground truth semantics", i.e., doctor analyze, all together to be usable in preparing machine learning calculations. Asking for this as extra contributions from officially over-burden doctors is, be that as it may, infeasible. Accordingly elective innovative strategies are required for producing the preparation information for our setting. An appealing plausibility here is the capacity to use clinical records, which are too ending up more promptly open through the sending of e-Health records frameworks. Figure 2 outlines this structure for information examination: the earlier sensor information with related information from the clinical records, mined crosswise over numerous people, can frame the reason for machine realizing where the doctor analyze that are as of now some portion of the clinical record give the essential semantic naming of ground truth once they are properly transiently lined up with the sensor information. The educated arrangement and relapse philosophies would then be able to be utilized with current information to advise the doctor's present forecasts/analyze. The procedure would then be able to be iterated. The benefit of making the linkage with the clinical records is that continuous clinical practice can give preparing information for the machine learning with next to zero extra weight on the doctors.

- Thirdly, the information contributions for deduction are profoundly heterogeneous. The tangible information speak to altogether different modalities. The statistic and verifiable data in the clinical records, albeit to a great degree instructive for deduction is generally of an altogether different nature from the sensor information. This scope of heterogeneity challenges regular machine learning approaches that bargain basically with homogeneous information. Graphical models [63] that permit blend of heterogeneous contributions to a basic structure are in this way prone to be useful for surmising in these settings; however these are additionally prone to require critical customization to be successful




Fig. 2. Analytics workflow for systems integrating wearable sensor technology into clinical practice.

VII. VISUALIZATION

It is unreasonable to request that doctors pore over the voluminous information or investigations from IoT-based sensors. To be helpful in clinical practice, the outcomes from the Analytics Engine require to be displayed to doctors in an instinctive arrangement where they can promptly fathom the between relations between amounts what's more, in the end begin utilizing the tangible information in their clinical practice. Representation is perceived as an autonomous and essential research region with a wide exhibit of utilizations in

both science and everyday life [64]. Given that shading is a key discriminative trait of our visual recognition, it is obvious that shading assumes a key job in data representation. Shading separation and shading class have been appeared be viable in permitting fast distinguishing proof and appreciation of contrasts in outwardly introduced information [65]. The kind of shading representation that is best is subject to the sort of information. For one-dimensional relations, standards for representation of relations have additionally been built up: Qualitative palettes, where the equivalent perceptual significance is given to all hues, are favored when clear cut information is transmitted, like reference diagrams, pie outlines, and so forth. Consecutive palettes, where hues with daintiness contrasts are allotted, are proposed for numerical factors whose esteem runs in an interim. For grouping maps, veering palettes, that blend the subjective what's more, consecutive palette directions, are considered as the best choice [66]. Representation of multi-dimensional information, on the other hand, utilizes a blend of shading, spatial area and different properties and remains a testing issue.

Information assembled or gathered from IoT sensors ranges the total range of classifications sketched out in the past passage furthermore, consequently a variety of various perception approaches are required for powerful utilization of the information. An unmistakable part of wearable sensor information, with respect to information procured at a lab or then again amid a clinical visit, is that the information are assembled over a any longer longitudinal span, with a better transient inspecting, furthermore, at the same time over different modalities. While the information speaks to a fortune trove for machine learning and derivation, for the doctor, it is tricky in the nonappearance of apparatuses to promptly picture and communicate with the information. The time-

differing and multi-dimensional parts of the information present a specific test in light of the fact that these have commonly not been utilized in clinical practice despite the fact that the worldly variety what's more, development of information and examinations results are of specific enthusiasm for conclusion.

To feature the utility of viable perception of worldly data, we present a solid case for cardio-vascular infection (CVD) observing [6], [67], [68]. Holter monitor based ECG recording over lengths of 24-48 hours is as of now used in CVD conclusion [69]. In addition to other things, such checking is valuable for recognizing anomalous extension of the QT interim, which speaks to the length of time taken for electrical depolarization and repolarization of the ventricles also, estimated on the ECG as the span between the beginning of the Q-wave and the finish of the T-wave [70]. An unusual prolongation of this interim, called Long QT Syndrome (LQTS) is an imperative marker of potential breaking down of the heart [70], [71]. For analysis, a rectified esteem, QTc, that makes up for the normal variety in QT interim with the pulse [72] is more straightforwardly enlightening than the crude QT values. The QTc interim is more often than not around 400 ms in a solid individual, and may go up to 500 ms or significantly higher with LQTS.

After a patient experiences a holter based chronicle session, QTc values are usually gotten from examinations of ECG information and accessible to the doctor for the length of the investigation (one esteem for each heart beat). A cardiologist that has 20 patients may approach a table containing yesterday's two million QTc values, which can obviously not be separately analyzed as crude information. In current practice, cardiologists ordinarily spot-check around 10 seconds of the patient's ECG, and audit the processed normal qualities over an entire 24-hour recording. This process disposes of a great deal of key data. On account of LQTS, for example, QTc could be drawn out for a few minutes or indeed, even hours without the specialist seeing the issue. There is a clinical requirement for a superior method to picture the full informational collection. Figure 3 delineates a plan that we have as of late created for picturing long haul checking of amended QT (QTc) results [6], [67], [68]. The two plots appeared in Figure 3 indicate 24 hour Holter checking results plotted inside a circle. Midnight is the highest point of the plot and twelve is the base. Low QTc values are inside the circle, and high qualities are along the edges. Diverse shading groups are the QTc edges for ordinary (green: 360– 425 ms), marginal (yellow: 450– 500 ms), and unusual (red: ≥ 500 ms). While the solid patient's QTc (top figure) dependably remains inside the typical area, the unfortunate patient's QTc (base figure) advances into the unusual area. These sort of plots appeared in Figure 3 permit for the doctor to promptly observe and grasp the full fleeting variety in QTc over the whole chronicle interim, rather than spotting check singular qualities. Note additionally the huge change in QTc during the evening in Figure 3, which can't be recognized in clinical ECG checking done amid the day.

While this model concentrated on QTc and 24-hour perception periods, the procedure and system will be like screen other therapeutic markers, for example, O₂ immersion or glucose levels, furthermore, over various time interims. The previous precedent represented the representation of one parameter over the transient span of the chronicle through an educational picture, which we note was static instead of




Fig. 3. QTc (in seconds) over 24 hours using the Bazett correction equation [73]. Top: healthy 24yo male patient. Bottom: 35yo male patient with the LQT2 genetic mutation, on beta blockers. “Slices” in the plots indicate a period that was not recorded. The green band is the interquartile range for healthy male patients in the THEW database [74]. Red represents abnormal and potentially dangerous QTc values fluctuating with time. One can likewise promptly present various parameters in parallel by means of one such picture for every parameter. While this can be proficient for the doctor to see the variety in every individual parameter initially, it isn't as instinctive for seeing how the showed parameters may co-shift after some time. To address, this, extra representation techniques are required for imagining the fleeting measurement. The intuitiveness accessible through touch interfaces in present day portable gadgets, for example, cell phones and tablets offers an especially alluring open door for representation of transient relations. We have just been utilizing such intuitiveness for making strides the entrance of shading insufficient people to shading symbolism [75]. By utilizing intuitiveness, this interface likewise enables us to imagine changes after some time, which as we officially noted, is a specific viewpoint that makes wearable sensor information especially valuable yet in addition testing to use in doctors clinical settings. While there has been earlier work on perception of time-related information in organic settings [76], the center has been totally on static pictures suited for incorporation in a distribution. Unique in relation to these, approaches especially adjusted for intelligent utilize that naturally enables the doctor to scroll forward and backward so as to evaluate the worldly development and coevolution of various amounts of enthusiasm utilizing cell phone what's more, tablet gadgets as of now being conveyed in their workplaces.

VIII.SUMMARY AND CONCLUDING REMARKS

In this paper, we explored the present state and anticipated future bearings for incorporation of remote wellbeing checking innovations into the clinical routine with regards to drug. Wearable sensors, especially those furnished with IoT knowledge, offer appealing choices for empowering perception and recording of information in home and workplaces, over any longer terms than are as of now done at office and research facility visits. This fortune trove of information, when investigated and introduced to doctors in simple to-acclimatize representations has the potential for profoundly enhancing human services and decreasing expenses. We featured a few of the difficulties in detecting, examination, and representation that should be tended to before frameworks can be intended for consistent mix into clinical practice.

IX. ACKNOWLEDGMENT

This work was upheld to some degree by the National Science Establishment allow CNS-1239423 and a blessing from Nvidia Corp.

REFERENCES

- [1] Jawbone Inc., “Jawbone fitness trackers,” accessed April 2015. [Online]. Available: <https://jawbone.com/up/trackers>
- [2] FitBit Inc., “flex: Wireless activity + sleep wristband,” accessed April 015. [Online]. Available: <https://www.fitbit.com/flex>
- [3] Apple Inc., “Apple watch,” accessed April 2015. [Online]. Available: <https://www.apple.com/watch/>
- [4] A. Pantelopoulos and N. Bourbakis, “A survey on wearable sensor-based systems for health monitoring and prognosis,” *IEEE Trans. Sys., Man, and Cybernetics, Part C: Applic. and Reviews*, vol. 40, no. 1, pp. 1–12, Jan 2010.
- [5] D. Son, J. Lee, S. Qiao, R. Ghaffari, J. Kim, J. E. Lee, C. Song, S. J. Kim, D. J. Lee, S. W. Jun, S. Yang, M. Park, J. Shin, K. Do, M. Lee, K. Kang, C. S. Hwang, N. Lu, T. Hyeon, , and D.-H. Kim, “Multifunctional wearable devices for diagnosis and therapy of movement disorders,” *Nature anotechnology*, pp. 1–8, 2014.
- [6] A. Page, O. Kocabas, T. Soyata, M. Aktas, and J.-P.Couderc, “Cloud-Based Privacy-Preserving Remote ECG Monitoring and Surveillance,” *Annals of Noninvasive Electrocardiology (ANEC)*, 2014. [Online]. Available: <http://dx.doi.org/10.1111/anec.12204>
- [7] R. Paradiso, G. Loriga, and N. Taccini, “A wearable health care system based on knitted integrated sensors,” *IEEE Trans. Info. Tech. in Biomedicine*, vol. 9, no. 3, pp. 337–344, Sept 2005.
- [8] A. Milenkovi, C. Otto, and E. Jovanov, “Wireless sensor networks for personal health monitoring: Issues and an implementation,” *Comput. Commun.*, vol. 29, no. 1314, pp. 2521 – 2533, 2006, wirellessSenson Networks and Wired/Wireless Internet Communications. [Online]. Available: <http://www.sciencedirect.com/science/article/pii/S0140366406000508>

- [9] M. Bazzani, D. Conzon, A. Scalera, M. Spirito, and C. Trainito, "Enabling the IoT paradigm in e-health solutions through the VIRTUS middleware," in *IEEE 11th Int. Conf. on Trust, Security and Privacy in Computing and Com. (TrustCom)*, June 2012, pp. 1954–1959.
- [10] A. Benharref and M. Serhani, "Novel cloud and SOA-based framework for E-Health monitoring using wireless biosensors," *IEEE Journal of Biomed. and Health Inf.*, vol. 18, no. 1, pp. 46–55, Jan 2014.
- [11] S. Babu, M. Chandini, P. Lavanya, K. Ganapathy, and V. Vaidehi, "Cloud-enabled remote health monitoring system," in *Int. Conf. on Recent Trends in Inform. Tech. (ICRTIT)*, July 2013, pp. 702–707.
- [12] C. Rolim, F. Koch, C. Westphall, J. Werner, A. Fracalossi, and G. Salvador, "A cloud computing solution for patient's data collection in health care institutions," in *Second Int. Conf. on eHealth, Telemedicine, and Social Medicine, ETELEMED '10.*, Feb 2010, pp. 95–99.
- [13] M. Lan, L. Samy, N. Alshurafa, M.-K. Suh, H. Ghasemzadeh, A. Macabasco-O'Connell, and M. Sarrafzadeh, "Wanda: An end-to-end remote health monitoring and analytics system for heart failure patients," in *Proc. of the Conf. on Wireless Health*, ser. WH '12. New York, NY, USA: ACM, 2012, pp. 9:1–9:8.
- [14] L. Wei, N. Kumar, V. Lolla, E. Keogh, S. Lonardi, C. Ratanamahatana, and H. Van Herle, "A practical tool for visualizing and data mining medical time series," in *Proc. 18th IEEE Symposium on Computer-Based Med. Sys.*, June 2005, pp. 341–346.
- [15] Y. Mao, Y. Chen, G. Hackmann, M. Chen, C. Lu, M. Kollef, and T. Bailey, "Medical data mining for early deterioration warning in general hospital wards," in *IEEE 11th Int. Conf. on Data Mining Workshops (ICDMW)*, Dec 2011, pp. 1042–1049.
- [16] V. Ukis, S. TirunellaiRajamani, B. Balachandran, and T. Friese, "Architecture of cloud-based advanced medical image visualization solution," in *IEEE Int. Conf. on Cloud Computing in Emerging Markets (CCEM)*, Oct 2013, pp. 1–5.
- [17] B. Rao, "The role of medical data analytics in reducing health fraud and improving clinical and financial outcomes," in *Computer-Based Medical Systems (CBMS), 2013 IEEE 26th International Symposium on*, June 2013, pp. 3–3.
- [18] P. Ray, "Home health hub internet of things (H3IoT): An architectural framework for monitoring health of elderly people," in *Int. Conf. on Science Eng. and Management Research (ICSEMR)*, Nov 2014, pp. 1– 3.
- [19] N. Bui and M. Zorzi, "Health care applications: A solution based on the internet of things," in *Proc. of the 4th Int. Symposium on Applied Sciences in Biomed. and Com. Tech.*, ser. ISABEL '11. New York, NY, USA: ACM, 2011, pp. 131:1–131:5.
- [20] W. Zhao, C. Wang, and Y. Nakahira, "Medical application on internet of things," in *IET Int. Conf. on Com. Tech. and Application (ICCTA 2011)*, Oct 2011, pp. 660–665.
- [21] F. Hu, D. Xie, and S. Shen, "On the application of the internet of things in the field of medical and health care," in *IEEE Int. Conf. on and IEEE Cyber, Physical and Social Computing Green Computing and Communications (GreenCom),(iThings/CPSCoM)*, Aug 2013, pp. 2053– 2058.
- [22] T. Soyata, R. Muraleedharan, C. Funai, M. Kwon, and W. Heinzelman, "Cloud-Vision: Real-Time face recognition using a Mobile-Cloudlet- Cloud acceleration architecture," in *Proceedings of the 17th IEEE Symposium on Computers and Communications (IEEE ISCC 2012)*, Cappadocia, Turkey, Jul 2012, pp. 59–66.

- [23] G. Nalinipriya and R. Aswin Kumar, "Extensive medical data storage with prominent symmetric algorithms on cloud - a protected framework," in *IEEE Int. Conf. on Smart Structures and Systems (ICSSS)*, March 2013, pp. 171–177.
- [24] A. F. M. Hani, I. V. Paputungan, M. F. Hassan, V. S. Asirvadani, and M. Daharus, "Development of private cloud storage for medical image research data," in *Int.Conf. on Computer and Inf. Sciences (ICCOINS)*, June 2014, pp. 1–6.
- [25] "World health organization factsheets: Cardiovascular diseases (CVDs)," online, accessed April 2015. [Online]. Available: <http://www.who.int/mediacentre/factsheets/fs317/en/>
- [26] "World health organization factsheets: Chronic obstructive pulmonary disease (COPD)," online, accessed April 2015. [Online]. Available: <http://www.who.int/mediacentre/factsheets/fs315/en/>
- [27] "Parkinson's disease information (mayo clinic website)," online, accessed April 2015. [Online]. Available: <http://www.mayoclinic.org/diseases-conditions/parkinsons-disease/basics/definition/con-20028488>
- [28] E. Dorsey, R. Constantinescu, J. Thompson, K. Biglan, R. Holloway, K. Kieburtz, F. Marshall, B. Ravina, G. Schifitto, A. Siderow *et al.*, "Projected number of people with Parkinson disease in the most populous nations, 2005 through 2030," *Neurology*, vol. 68, no. 5, pp. 384–386, 2007.
- [29] S. Xu, Y. Zhang, L. Jia, K. E. Mathewson, K.-I. Jang, J. Kim, H. Fu, X. Huang, P. Chava, R. Wang, S. Bhole, L. Wang, Y. J. Na, Y. Guan, M. Flavin, Z. Han, Y. Huang, and J. A. Rogers, "Soft microfluidic assemblies of sensors, circuits, and radios for the skin," *Science*, vol. 344, pp. 70–74, 2014.
- [30] D.-H. Kim, R. Ghaffari, N. Lu, and J. A. Rogers, "Flexible and stretchable electronics for biointegrated devices," *Annual Review of Biomedical Engineering*, pp. 113–128, 2012.
- [31] O. Olorode and M. Nourani, "Reducing leakage power in wearable medical devices using memory nap controller," in *Circuits and Sys. Conf. (DCAS), IEEE Dallas*, Oct 2014, pp. 1–4.
- [32] C. Park, P. Chou, Y. Bai, R. Matthews, and A. Hibbs, "An ultra-wearable, wireless, low power eeg monitoring system," in *Biomed. Circuits and Sys. Conf., BioCAS 2006. IEEE*, Nov 2006, pp. 241–244.
- [33] T. Torfs, V. Leonov, C. Van Hoof, and B. Gyselinckx, "Body-heat powered autonomous pulse oximeter," in *5th IEEE Conf. on Sensors*, Oct 2006, pp. 427–430.
- [34] S. Madhani, M. Tauil, and T. Zhang, "Collaborative sensing using uncontrolled mobile devices," in *Int. Conf. on Collaborative Computing: Networking, Applications and Worksharing*, 2005, pp. 8 pp.–.
- [35] C. T. Chou, R. Rana, and W. Hu, "Energy efficient information collection in wireless sensor networks using adaptive compressive sensing," in *IEEE 34th Conf. on Local Computer Networks, LCN 2009.*, Oct 2009, pp. 443–450.
- [36] H. Zhang and J. Hou, "Maintaining sensing coverage and connectivity in large sensor networks," *Ad Hoc & Sensor Wireless Networks*, vol. 1, no. 1-2, pp. 89–123, 2005.
- [37] C. Huang and Y. Tseng, "The Coverage Problem in a Wireless Sensor Network," *Mobile Networks and Applications*, vol. 10, no. 4, pp. 519–528, 2005.
- [38] Y. Chen and Q. Zhao, "On the lifetime of wireless sensor networks," *IEEE Commun.Lett.*, vol. 9, no. 11, pp. 976–978, 2005.

- [39] M. Cardei and J. Wu, "Energy-efficient coverage problems in wireless ad-hoc sensor networks," *Computer Communications*, vol. 29, no. 4, pp. 413–420, 2006.
- [40] C. Yu and G. Sharma, "Camera scheduling and energy allocation for lifetime maximization in user-centric visual sensor networks," *IEEE Trans. Image Proc.*, vol. 19, no. 8, pp. 2042–2055, Aug. 2010.
- [41] J.-S. Lee, Y.-W. Su, and C.-C. Shen, "A comparative study of wireless protocols: Bluetooth, uwb, zigbee, and wi-fi," in *33rd Annual Conf. of the IEEE Indust. Elect. Society, IECON 2007.*, Nov 2007, pp. 46–51.
- [42] M. Siekkinen, M. Hienkari, J. Nurminen, and J. Nieminen, "How low energy is bluetooth low energy? comparative measurements with zigbee/802.15.4," in *Wireless Communications and Networking Conference Workshops (WCNCW), 2012 IEEE*, April 2012, pp. 232–237.
- [43] Y. Li and W. Wang, "The unheralded power of cloudlet computing in the vicinity of mobile devices," in *IEEE Globecom Workshops (GC Wkshps)*, Dec 2013, pp. 4994–4999.
- [44] M. Kwon, Z. Dou, W. Heinzelman, T. Soyata, H. Ba, and J. Shi, "Use of network latency profiling and redundancy for cloud server selection," in *Proceedings of the 7th IEEE International Conference on Cloud Computing (IEEE CLOUD 2014)*, Alaska, USA, Jun 2014, pp. 826–832.
- [45] T. Soyata, R. Muraleedharan, S. Ames, J. H. Langdon, C. Funai, M. Kwon, and W. B. Heinzelman, "Combat: mobile cloud-based compute/communications infrastructure for battlefield applications," in *Proceedings of SPIE*, vol. 8403, May 2012, pp. 84 030K–84 030K.
- [46] M. Quwaider and Y. Jararweh, "Cloudlet-based for big data collection in body area networks," in *8th International Conference for Internet Technology and Secured Transactions (ICITST)*, Dec 2013, pp. 137–141.
- [47] Y. Jararweh, L. Tawalbeh, F. Ababneh, and F. Dosari, "Resource efficient mobile computing using cloudlet infrastructure," in *IEEE Ninth International Conference on Mobile Ad-hoc and Sensor Networks (MSN)*, Dec 2013, pp. 373–377.
- [48] T. Soyata, H. Ba, W. Heinzelman, M. Kwon, and J. Shi, "Accelerating mobile cloud computing: A survey," in *Communication Infrastructures for Cloud Computing*, H. T. Mouftah and B. Kantarci, Eds. Hershey PA, USA: IGI Global, Sep 2013, ch. 8, pp. 175–197.
- [49] A. Abbas and S. Khan, "A review on the state-of-the-art privacy-preserving approaches in the e-health clouds," *IEEE Journal of Biomedical and Health Informatics*, vol. 18, no. 4, pp. 1431–1441, July 2014.
- [50] National Institute of Standards and Technology, "Advanced encryption standard (AES)," Nov. 2001, FIPS-197.
- [51] M. Li, S. Yu, Y. Zheng, K. Ren, and W. Lou, "Scalable and secure sharing of personal health records in cloud computing using attribute-based encryption," *IEEE Trans. Parallel Distrib. Syst.*, vol. 24, no. 1, pp. 131–143, Jan. 2013.
- [52] S. Ruj, M. Stojmenovic, and A. Nayak, "Privacy preserving access control with authentication for securing data in clouds," in *Proceedings of the 2012 12th IEEE/ACM International Symposium on Cluster, Cloud and Grid Computing (Ccggrid 2012)*, ser. CCGRID '12. Washington, DC, USA: IEEE Computer Society, 2012, pp. 556–563.

- [53] M. Li, S. Yu, Y. Zheng, K. Ren, and W. Lou, "Scalable and secure sharing of personal health records in cloud computing using attribute-based encryption," *IEEE Transactions on Parallel and Distributed Systems*, vol. 24, no. 1, pp. 131–143, 2013.
- [54] C. Bishop, *Pattern recognition and machine learning*. New York, NY: Springer, 2006.
- [55] Mathworks Inc., *MATLAB The Language of Technical Computing*, Natick, Massachusetts, accessed April 2015. [Online]. Available: <http://www.mathworks.com/products/matlab/>
- [56] R Core Team, *R: A Language and Environment for Statistical Computing*, R Foundation for Statistical Computing, Vienna, Austria, accessed April 2015. [Online]. Available: <http://www.R-project.org/>
- [57] S. Shalev-Shwartz, Y. Singer, N. Srebro, and A. Cotter, "Pegasos: Primal estimated sub-gradient solver for SVM," *Mathematical Programming*, vol. 127, no. 1, pp. 3–30, 2011.
- [58] A. Bordes, S. Ertekin, J. Weston, and L. Bottou, "Fast kernel classifiers with online and active learning," *J. of Machine Learning Research*, vol. 6, pp. 1579–1619, 2005.
- [59] J. C. Platt, "Fast training of support vector machines using sequential minimal optimization," *Advances in Kernel Methods*, pp. pp. 185–208, 1999.
- [60] G. Chechik, G. Heitz, G. Elidan, P. Abbeel, and D. Koller, "Max-margin classification of data with absent features," *J. of Machine Learning Research*, vol. 9, pp. 1–21, 2008.
- [61] P. J. Garcia-Laencina, J. L. Sancho-Gomez, and A. R. Figueiras-Vidal, "Pattern classification with missing data: A review," *J. of Machine Learning Research*, vol. 19, no. 2, pp. 263–282, 2010.
- [62] R. J. A. Little and D. B. Rubin, *Statistical Analysis with Missing Data*. New Jersey: Wiley, 2002.
- [63] M. Jordan and T. Sejnowski, Eds., *Graphical models: foundations of neural computation*, ser. Computational neuroscience. MIT Press, 2001.
- [64] E. R. Tufte and P. Graves-Morris, *The visual display of quantitative information*. Graphics press Cheshire, CT, 1983, vol. 2.
- [65] C. Healey, "Choosing effective colours for data visualization," in *Proceedings of the 7th conference on Visualization*. IEEE, 1996, pp. 263–270.
- [66] L. D. Bergman, B. E. Rogowitz, and L. A. Treinish, "A rule-based tool for assisting colormap selection," in *Proceedings of the 6th conference on Visualization*. Washington, DC, USA: IEEE Computer Society, 1995, pp. 118–.
- [67] Mehmet Aktas and Alex Page and Tolga Soyata and JP Couderc, "Incidence of diurnal QT concealment in genotype positive long QT1 and 2 patients," in *Heart Rhythm Society's Annual Scientific Sessions*, 2015, to be presented, May 13-16, 2015, Boston, MA.
- [68] O. Kocabas, T. Soyata, J.-P. Couderc, M. Aktas, J. Xia, and M. Huang, "Assessment of cloud-based health monitoring using homomorphic encryption," in *Proceedings of the 31st IEEE International Conference on Computer Design (ICCD)*, Ashville, VA, USA, Oct 2013, pp. 443–446.
- [69] "Holter monitor," online, accessed April 2015. [Online]. Available: <http://www.nlm.nih.gov/medlineplus/ency/article/003877.htm>
- [70] "Long QT syndrome (mayo clinic website)," online, accessed April 2015. [Online]. Available: <http://www.mayoclinic.org/diseases-conditions/long-qt-syndrome/basics/causes/con-20025388>

[71] H. Morita, J. Wu, and D. P. Zipes, "The QT syndromes: long and short," *The Lancet*, vol. 372, no. 9640, pp. 750 – 763, 2008. [Online]. Available:

<http://www.sciencedirect.com/science/article/pii/S0140673608613070>

[72] L. S. Fridericia, "Die SystolendauerimElektrokardiogrammbeinormalenMenschen und beiHerzkranken," *ActaMedicaScandinavica*, vol. 53, pp. 469–486, 1920.

[73] H. C. Bazett, "An Analysis of Time Relations of the Electrocardiogram," *Heart*, vol. 7, pp. 353–370, 1920.